

## LAMPIRAN 1

### Hasil percobaan uji validitas pemeriksaan laju endap darah (LED) dengan metode Westergren, metode CLSI 2011, dan metode rujukan ICSH 1993 pasca 1 dan 2 jam

No.	ICSH 1993		CLSI 2011		Westergren	
	1 jam	2 jam	1 jam	2 jam	1 jam	2 jam
1	5	20	1	10	5	20
2	4	25	10	24	8	20
3	2	13	1	10	1	11
4	4	10	10	20	3	9
5	8	19	12	21	10	20
6	17	43	18	40	14	28
7	15	19	11	24	1	1
8	5	8	6	21	5	18
9	42	80	35	65	25	50
10	5	8	22	38	30	20
11	5	27	38	65	1	1
12	4	15	5	8	3	3
13	5	20	7	20	7	15
14	4	15	5	15	5	12
15	5	20	8	16	9	24
16	4	15	5	14	5	12
17	25	40	15	25	1	1
18	5	10	7	12	7	11
19	1	1	1	1	1	1
20	4	9	4	11	1	1
21	5	10	3	7	8	15
22	5	10	8	16	5	10
23	6	14	10	22	9	20
24	18	40	10	21	10	19
25	20	44	19	40	3	7
26	3	10	8	14	9	16

27	30	59	31	60	16	30
28	30	60	24	44	13	27
29	2	7	5	10	3	9
30	25	52	27	54	12	25

---

**LAMPIRAN 2****TABEL HASIL UJI *POST HOC FISHER* LSD***Dependent Variable: LED 2 Jam*

LSD

(I) Metode- 2 Jam	(J) Metode- 2 Jam	Mean Difference (I-J)	Std. Error	Sig.	95% Confidence Interval	
					Lower Bound	Upper Bound
CLSI	Na Citrate	9.87(*)	4.513	.031	.90	18.84
	ICSH	-1.93	4.513	.669	-10.90	7.04
Na Citrate	CLSI	-9.87(*)	4.513	.031	-18.84	-.90
	ICSH	-11.80(*)	4.513	.011	-20.77	-2.83
ICSH	CLSI	1.93	4.513	.669	-7.04	10.90
	Na Citrate	11.80(*)	4.513	.011	2.83	20.77

\* The mean difference is significant at the .05 level.